



To: All Potential Bidders
From: Anna Silva, Procurement Specialist
Date: March 26, 2019
Subject: IFB B79-19, Skull Clamps, Addendum #1

Please add this Addendum #1 to the original bid documents. It shall be the responsibility of the interested bidders to adhere to any changes or revision to the IFB as identified in this Addendum #1.

The following clarification shall become permanent and made part of the procurement file.

- 1) "Mayfield Infinity" and "Doro QR3" are two different manufactures and have different specs. Which one should we be more true to?

Bids should be "equal to or better than" either of these systems. Both systems are considered to be clinically acceptable.

- 2) "360" – This could mean several things to the potential bidders. Can you let us know which of these you mean? (see below for images.)
 - a. 360 degree swivel adapters.
 - b. Ball and socket swivel adapters.
 - c. 360 degree swiveling base units

Equipment must provide 360 degree position of patients' skull.

If there are any questions or inquiries in relation to this Addendum #1, Bidders may contact Anna Silva at (505) 272-9571 or by email at amsilva@salud.unm.edu.